


OXYGEN - oxygen gas
Community Medical Equipment


Disclaimer: This drug has not been found by FDA to be safe and effective, and this labeling has not been approved by FDA. For further information about unapproved drugs, click [here](#).

Oxygen

Principal for Oxygen Product



**OXYGEN
COMPRESSED
U.S.P
UN 1072**



Transfilled by:
Community Medical Equipment
107 Market St.
Glasgow, MO 65254
(660) 338-9965

Use: For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, Rx Only.

PRODUCED BY AIR LIQUEFACTION

WARNING

- Do not use high concentrations for more than 5 hours without 1-hour interruption. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment.
- Keep cylinder, cylinder valve, all equipment and connections free of oil and grease to avoid violent ignition.
- Do not use or store near heat or open flame.
- Use only with pressure reducing equipment and apparatus designed for use with oxygen.
- Federal law requires that this container be refilled with oxygen USP only by establishments registered as a drug producer in accordance with the federal food, drug and cosmetic act.

9217

LOT #
DATE

10

13

09

Oxygen Refrigerated Liquid U.S.P.

UN1073

WARNING!

High Pressure Oxidizing Liquid and Gas Under Pressure. Vigorously Accelerates Combustion. May Explode on Ignition or Impact. Can cause Severe Frostbite.

PRODUCED BY AIR LIQUEFACTION



Warning! Always Keep Container In Upright Position.

Rx Only

Keep oil, grease, and combustibles away. Use only with equipment cleared for oxygen service. Store and use with adequate ventilation. Do not get liquid in eyes, on skin or clothing. For liquid withdrawal, wear face shield and gloves. Do not drop. Use hand truck for container movement. Avoid spills. Do not walk on or roll equipment over spills. Close valve after each use and when empty.

In Case of Frostbite, obtain Medical treatment IMMEDIATELY.

Contents under pressure and can vigorously accelerate combustion. Keep free from oil and grease. Do not use or store near heat or open flame and use only with equipment conditioned for oxygen service.

Federal law requires that this container be refilled with Liquid Oxygen, U.S.P. only by establishments registered as drug producers in accordance with the Federal Food, Drug and Cosmetic Act.

Transfilled by: Community Medical Equipment • 107 Market St.
Glasgow, MO 65254 • (660) 338-9965

LOT# _____ EXP. _____ CONTENTS _____

DO NOT REMOVE THIS LABEL

OXYGEN

oxygen gas

Product Information

Product Type	HUMAN PRESCRIPTION DRUG	Item Code (Source)	NDC:63983-0001
Route of Administration	RESPIRATORY (INHALATION)		

Active Ingredient/Active Moiety

Ingredient Name

Basis of Strength

Strength

Oxygen (UNII: S88TT14065) (Oxygen - UNII:S88TT14065)	Oxygen	99 L in 100 L
--	--------	---------------

Packaging

#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NDC:63983-0001-1	170 L in 1 CYLINDER		
2	NDC:63983-0001-2	425 L in 1 CYLINDER		
3	NDC:63983-0001-3	679 L in 1 CYLINDER		
4	NDC:63983-0001-4	46 L in 1 DEWAR		

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
unapproved medical gas		02/13/1998	

Labeler - Community Medical Equipment (364293761)

Establishment

Name	Address	ID/FEI	Business Operations
Community Medical Equipment		364293761	manufacture

Revised: 10/2009

Community Medical Equipment