

**HONEYWELL STERILE EYESALINE EMERGENCY EYEWASH- water liquid**  
**Sperian Eye & Face Protection Inc**

*Disclaimer: Most OTC drugs are not reviewed and approved by FDA, however they may be marketed if they comply with applicable regulations and policies. FDA has not evaluated whether this product complies.*

-----  
**Sperian Eyesaline Emergency Eyewash**

***Drug Facts***

***Active Ingredient***

Sterile water 99%

***Purpose***

Emergency Eyewash

***Uses***

For flushing the eye to reduce chances of severe injury caused by acid, alkali, or particulate contamination.

***Warnings***

**For external use only**

**Do not use**

- for injection
- in intraocular surgery
- internally

**Ask a doctor before use if you have**

- eye pain
- changes in vision
- redness or irritation of the eye after use
- an injury caused by alkali

**Keep out of reach of children.** If swallowed, get medical help or contact a Poison Control Center right away.

***Directions***

- do not dilute solution or reuse cartridge
- pull handle on the activating door to initiate flow
- lower head into flow for a minimum of 20 minutes
- continue flushing with water if necessary
- obtain medical treatment

***Other information***

- store at room temperature 59° to 86° F(15° to 30°C)
- do not freeze

***Inactive ingredients***

sodium chloride, sodium phosphate dibasic, sodium phosphate monobasic

***Questions?***

call **1-800 - 430-5490**

Sperian Eye and Face Protection, Inc. (a Honeywell Company)

825 East Highway 151  
Platteville, WI 53818

## WARNING!

### FOR PROPER OPERATION:

1. Read all warnings and the Operation and Installation Manual prior to installation of cartridges. Failure to follow this installation instructions and warnings may result in personal injury, serious illness, including blindness or death.
2. Inspect cartridges for damage: If the nozzle assembly, flexible solution bag and carton has been damaged, or the nozzle assembly or bag is leaking fluid, DO NOT INSTALL! Call 1-800-430-5490.
3. The top of the carton must be unobstructed and free from damaged.
4. Install cartridges (two required) in the Fendall Pure Flow 1000 with slot A to the rear and then nozzle assembly in front.
5. Each cartridge weighs 32 lbs (14.5 kg). Use caution when handling and installing.
6. Cartridges must be replaced after use and on, or before the solution expiration date printed on the carton and activating strap.

## Honeywell STERILE eyesaline®

Emergency Eyewash

Buffered Isotonic Solution

3.9 Gallons (14.4L)

Net Wt. 32 lbs (14.5kg)

## #32-ST1050-0000



Emergency Eyewash  
Lavaojos de emergencia  
Lavage oculaire d'urgence

Buffered Isotonic Solution  
Solución Isotónica Estabilizadora  
Solution Isotonique Tamponée

Cartridge for /  
Cartucho para /  
Cartouche pour /  
Fendall Pure Flow 1000®

3.8 Gallons (14.4L)

Net Wt. 32 lbs (14.5kg)

 <p><b>FRONT / FRENTE / AVANT</b></p>	 <p><b>SLOT A / RAMURA A / FENTE A</b></p> <p><b>BACK / PARTE POSTERIOR / ARRIERE</b></p>	<p><b>⚠ WARNING!</b> <b>FOR PROPER OPERATION:</b></p> <ol style="list-style-type: none"><li>1) Read all warnings and the Operation and Installation Manual prior to installation of cartridges. Failure to follow this installation instructions and warnings may result in personal injury, serious illness, including blindness or death.</li><li>2) Inspect cartridges for damage: If the nozzle assembly, flexible solution bag and carton has been damaged, or the nozzle assembly or bag is leaking fluid, DO NOT INSTALL! Call 1-800-430-5490.</li><li>3) The top of the carton must be unobstructed and free from damaged.</li><li>4) Install cartridges (two required) in the Fendall Pure Flow 1000 with slot A to the rear and the nozzle assembly in front.</li><li>5) Each cartridge weighs 32 lbs. (14.5kg). Use caution when handling and installing.</li><li>6) Cartridges must be replaced after use and on, or before the solution expiration date printed on the carton and activating strap.</li></ol>	<p><b>⚠ ¡ADVERTENCIA!</b> <b>PARA EL FUNCIONAMIENTO CORRECTO:</b></p> <ol style="list-style-type: none"><li>1) Lea todas las advertencias y el manual de instalación y funcionamiento antes de instalar los cartuchos. El no seguir estas instrucciones de instalación ni las advertencias podría ocasionar serias lesiones personales, enfermedades serias, inclusive ceguera o la muerte.</li><li>2) Inspeccione los cartuchos: Si el ensamblaje de la boquilla, la bolsa flexible de solución o la caja han sido dañados, o hay fugas de líquido en el ensamblaje de la boquilla o bolsa, NO LOS INSTALE! Llame al 1-800-430-5490.</li><li>3) La parte superior de la caja y la ranura en la parte posterior de la caja debe estar libre de obstrucciones y de daños.</li><li>4) Instale los cartuchos (se necesitan dos) en la unidad Fendall Pure Flow 1000 con la ranura "A" hacia la parte posterior y el ensamblaje de la boquilla en el frente.</li><li>5) Cada cartucho pesa 14.5kg (32 libras). Tenga cuidado al manejarlos e instalarlos.</li><li>6) Los cartuchos deben reemplazarse después de su uso y on, o antes de la fecha de expiración de solución impresa en la tira de activación.</li></ol>	<p><b>⚠ AVERTISSEMENT!</b> <b>POUR ASSURER UN BON FONCTIONNEMENT:</b></p> <ol style="list-style-type: none"><li>1) Lire tous les avertissements et le Guide d'Installation et de fonctionnement, avant d'installer les cartouches. Ne pas suivre ces instructions d'installation ni ces avertissements risque de provoquer de graves blessures corporelles, des maladies graves voire la cécité ou la mort.</li><li>2) Inspecter les cartouches pour vérifier s'il n'y a pas de dommage. Si les gicleurs, le sac souple de solution ou le carton sont endommagés, ou si les gicleurs ou le sac présentent une fuite, NE PAS LES INSTALLER! Téléphoner sans frais au: 1-888-212-7233.</li><li>3) S'assurer que le dessus du carton et la fente située à l'arrière de ce dernier sont dégagés et intactes. Installer les cartouches (deux sont requises) dans le bassin Fendall Pure Flow 1000, en s'assurant d'orienter la fente à l'arrière et les gicleurs vers l'avant.</li><li>4) Chaque cartouche pèse 14,5 kg (32 livres). Manipuler et installer les cartouches avec soin.</li><li>5) Remplacer les cartouches après utilisation, et avant la date d'expiration de la solution imprimée sur la courroie noire.</li></ol>
		<p>US Pat. 5566406, 5695124, 5850641 CDM Pat. 2571661, 2575966</p>		

## HONEYWELL STERILE EYESALINE EMERGENCY EYEWASH

water liquid

**Product Information**

<b>Product Type</b>	HUMAN OTC DRUG	<b>Item Code (Source)</b>	NDC:64809-103
<b>Route of Administration</b>	OPHTHALMIC		

**Active Ingredient/Active Moiety**

<b>Ingredient Name</b>	<b>Basis of Strength</b>	<b>Strength</b>
WATER (UNII: 059QF0KO0R) (WATER - UNII:059QF0KO0R)	WATER	990 mL in 1 L

**Inactive Ingredients**

<b>Ingredient Name</b>	<b>Strength</b>
SODIUM PHOSPHATE, DIBASIC (UNII: GR686LBA74)	
SODIUM CHLORIDE (UNII: 451W47IQ8X)	
SODIUM PHOSPHATE, MONOBASIC (UNII: 3980JIH2SW)	

**Packaging**

#	<b>Item Code</b>	<b>Package Description</b>	<b>Marketing Start Date</b>	<b>Marketing End Date</b>
1	NDC:64809-103-14	14.4 L in 1 CARTRIDGE; Type 0: Not a Combination Product	06/18/2018	

**Marketing Information**

<b>Marketing Category</b>	<b>Application Number or Monograph Citation</b>	<b>Marketing Start Date</b>	<b>Marketing End Date</b>
OTC monograph not final	part349	06/18/2018	

**Labeler** - Sperian Eye & Face Protection Inc (013435034)**Establishment**

<b>Name</b>	<b>Address</b>	<b>ID/FEI</b>	<b>Business Operations</b>
Sperian Eye & Face Protection Inc		013435034	manufacture(64809-103)

Revised: 6/2018

Sperian Eye &amp; Face Protection Inc