$50\ TD\ GOLD\ GLUCOSE\ CONTROL\ SOLUTION\ LEVEL\ 1\ SET-$  glucose dehydrogenase, glucose

EPS Bio Technology Corp.

-----

## Home Use Medical Device- 50 TD Gold Glucose Control Solution Level 1

### 50 TD Gold Glucose Control Solution Level 1 Box



TD Gold Glucose Control Solution Level 1 label

# SIZE: 55 X 30 mm



Recap the bottle immediately after use. Discard test strips in 90 days after first opening date.

Open on /

LOT

**EXP** 

- TD Gold Glucose Control Solution is for in vitro diagnostic use only.
- Rock the control solution bottle back and forth several times before use.
- DO NOT freeze.
- Use the control solution only for testing with TD Gold Meters and Test Strips.
- Store the TD Gold Glucose Control Solution between 36-86°F (2-30°C).

Manufactured for Total Diabetes, LLC Telephone Assistance / 855-525-8332

Manufacturer / EPS Bio Technology Corp. PN 706400015\_1248\_02

# 50 TD GOLD GLUCOSE CONTROL SOLUTION LEVEL 1 SET

glucose dehydrogenase, glucose

#### **Product Information**

Product Type MEDICAL DEVICE Item Code (Source) NHRIC:47884-189

#### **Product Characteristics**

number of times usable 4000

#### **Packaging**

7	# Item Code	Package Description	Marketing Start Date	Marketing End Date
	NHRIC:47884-189-50	50 in 1 PACKAGE		

**Marketing Information** 

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
premarket notification	K10 10 37	10/01/2013	

Revised: 10/2013 EPS Bio Technology Corp.