

**TD GOLD GLUCOSE CONTROL SOLUTION LEVEL 3- glucose dehydrogenase, glucose
EPS Bio Technology Corp.**

Home Use Medical Device:TD Gold Glucose Control Solution Level 3

TD Gold Glucose Control Solution Level 3

TD Gold
Glucose Control Solutions
Level 3 **4ml**

Recap the bottle immediately after use.
Discard test strips in 90 days after first opening date.

Open on /

LOT
EXP

- TD Gold Glucose Control Solution is for in vitro diagnostic use only.
- Rock the control solution bottle back and forth several times before use.
- DO NOT freeze.
- Use the control solution only for testing with TD Gold Meters and Test Strips.
- Store the TD Gold Glucose Control Solution between 36-86°F (2-30°C).

Manufactured for Total Diabetes, LLC
Telephone Assistance / 855-525-8332
NDC 47884-0180-41
Manufacturer / EPS Bio Technology Corp.
PN 706400020_1250_01

TD Gold Glucose Control Solution Level 3 Box



TD GOLD GLUCOSE CONTROL SOLUTION LEVEL 3

glucose dehydrogenase, glucose

Product Information

Product Type	MEDICAL DEVICE	Item Code (Source)	NHRIC:47884-180
---------------------	----------------	---------------------------	-----------------

Product Characteristics

number of times usable	80
-------------------------------	----

Packaging

#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NHRIC:47884-180-41	1 in 1 PACKAGE		

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
--------------------	--	----------------------	--------------------

premarket notification	K10 1037	08/30/2013	
------------------------	----------	------------	--

Labeler - EPS Bio Technology Corp. (658620117)

Registrant - EPS Bio Technology Corp. (658620117)

Revised: 9/2013

EPS Bio Technology Corp.