



March 21, 2024

DePuy Synthes
Jeffrey Krawiec
Senior Regulatory Affairs Program Lead
1301 Goshen Parkway
West Chester, Pennsylvania 19380

Re: K233665

Trade/Device Name: DePuy Synthes VOLT Mini Fragment Plating System, DePuy Synthes VOLT Small Fragment Plating System

Regulation Number: 21 CFR 888.3030

Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories

Regulatory Class: Class II

Product Code: HRS, HWC

Dated: February 21, 2024

Received: February 21, 2024

Dear Jeffrey Krawiec:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K233665

Device Name

DePuy Synthes VOLT™ Mini Fragment Plating System
DePuy Synthes VOLT™ Small Fragment Plating System

Indications for Use (Describe)

DePuy Synthes VOLT™ Mini Fragment Plating System:

The VOLT™ Mini Fragment Plating System is indicated for internal fracture fixation of bones and bone fragments of the appendicular skeleton appropriate for the implant size.

The VOLT™ Mini Fragment Plating System is intended for adults and both children (2-12 years) and adolescents (12-21 years) in which growth plates (physes) have fused or in which unfused growth plates will not be compromised by fixation.

If used in the femur, tibia, humerus, patella, or pelvis the VOLT™ Mini Fragment Plating System can only be used for non-load bearing stabilization and reduction.

DePuy Synthes VOLT™ Small Fragment Plating System:

The VOLT Small Fragment Plating System is indicated for internal fracture fixation of bones and bone fragments of the appendicular skeleton appropriate for the implant size.

The VOLT Small Fragment Plating System is intended for adults and both children (2-12 years) and adolescents (12-21 years) in which growth plates (physes) have fused or in which unfused growth plates will not be compromised by fixation.

If used in the femur the VOLT Small Fragment Plating System can only be used for non-load bearing stabilization and reduction.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(K) SUMMARY

Sponsor	Synthes (USA) Products, LLC 1301 Goshen Parkway West Chester, PA 19380
Contact	Jeffrey Krawiec, PhD Senior Regulatory Affairs Program Lead T: +1 484 816 6858 E: jkrawiec@its.jnj.com
Alternate Contact	Kathryn Schlesinger Regulatory Affairs Specialist II T: +41 76 503 78 43 E: kschlesi@its.jnj.com
Date Prepared	November 15, 2023
Proprietary Name	DePuy Synthes VOLT™ Mini Fragment Plating System DePuy Synthes VOLT™ Small Fragment Plating System
Classification Name	Single/multiple component metallic bone fixation appliances and accessories; Smooth or threaded metallic bone fixation fastener
Classification	Class II Regulation Numbers: 21 CFR §888.3030; 21 CFR §888.3040 Product Codes: HRS; HWC
Predicate Device	As this is a bundled traditional 510(k) submission and therefore multiple primary predicates are identified below. Primary predicates: <ul style="list-style-type: none"> • K000684 - Small Fragment Dynamic Compression Locking (DCL) System • K082807 - Synthes (USA) 3.5 and 4.5 mm Locking Compression Plate System with Expanded Indications • K011335 - Synthes One--Third Tubular DCL Plate • K033805 - Synthes (USA) 3.5 / 4.5 mm LCP Metaphyseal Plates • K131186 - Synthes 2.7mm and 3.5mm Variable Angle LCP Midfoot/Hindfoot System • Pre-amendment - Synthes 4.0mm Cancellous Screws • K063049 - Synthes (USA) Modular Mini Fragment LCP System • K082072 - Synthes (USA) 3.5mm LCP Hook Plate • K092247 - Synthes Locking Hand Plates • K100776 - Synthes 2.4mm/2.7mm Variable Angle (VA)-LCP Forefoot/Midfoot System • K112583 - Synthes Cortical Screws • K150099 - Depuy Synthes Variable Angle Locking Hand System (1.3mm and 2.0mm Plates and Screws)
Device Description	<u>DePuy Synthes VOLT™ Mini Fragment Plating System</u> The DePuy Synthes VOLT™ Mini Fragment Plating System is a family of implantable devices, consisting of non-contoured, non-anatomic straight and shaped plates (Adaption, Adaption Combi, Compact Straight, Straight,

	<p>Condylar, Hook, T-Plate, Tine, Y-Plate and Triangle), with variable angle screw holes, and two screw types (cortex and locking screws). The system is available in three sizes; 2.0mm, 2.4mm and 2.7mm. The Triangle Plate is available in 2.4mm and 2.7mm sizes only.</p> <p>The plates of this system are available in Stainless Steel and Commercially Pure Titanium, whilst the corresponding screws are available in Stainless Steel and Titanium Alloy (TAV) respectively. Plates and screws within the VOLT™ Mini Fragment Plating System are available either sterile or non-sterile and are single-use only.</p> <p><u>DePuy Synthes VOLT™ Small Fragment Plating System</u></p> <p>The DePuy Synthes VOLT™ Small Fragment Plating System is a family of implantable devices consisting of 3.5mm non-contoured, non-anatomic plates with variable angle screw holes. and 3.5mm locking, 3.5mm cortex, and 4.0mm cancellous screws.</p> <p>The plates of this system are available in Stainless Steel and Titanium, whilst the corresponding screws are available in Stainless Steel and Titanium Alloy (TAV). Plates and screws within the VOLT™ Small Fragment Plating System are available either sterile or non-sterile and are single-use only</p>
Indications for Use	<p><u>DePuy Synthes VOLT™ Mini Fragment Plating System</u></p> <p>The VOLT™ Mini Fragment Plating System is indicated for internal fracture fixation of bones and bone fragments of the appendicular skeleton appropriate for the implant size.</p> <p>The VOLT™ Mini Fragment Plating System is intended for adults and both children (2-12 years) and adolescents (12-21 years) in which growth plates (physes) have fused or in which unfused growth plates will not be compromised by fixation.</p> <p>If used in the femur, tibia, humerus, patella, or pelvis the VOLT™ Mini Fragment Plating System can only be used for non-load bearing stabilization and reduction.</p> <p><u>DePuy Synthes VOLT™ Small Fragment Plating System</u></p> <p>The VOLT™ Small Fragment Plating System is indicated for internal fracture fixation of bones and bone fragments of the appendicular skeleton appropriate for the implant size.</p> <p>The VOLT™ Small Fragment Plating System is intended for adults and both children (2-12 years) and adolescents (12-21 years) in which growth plates (physes) have fused or in which unfused growth plates will not be compromised by fixation.</p>

	If used in the femur the VOLT™ Small Fragment Plating System can only be used for non-load bearing stabilization and reduction.
Contraindications	No contraindications specific to these devices.
Non-Clinical Performance Testing	<p>To demonstrate the safety and efficacy of the subject devices and support the substantial equivalence to their predicates, the following testing was performed:</p> <ul style="list-style-type: none"> • Computational Finite Element Analysis simulating a mechanical static bending construct test, demonstrated subject plates were non-inferior regarding their bending moment and stiffness. • Engineering analyses and utilization of the Chapman Equation demonstrated subject screws were non-inferior screw regarding their torsional strength and axial pullout strength. • Assessment in accordance with the FDA Safety and Performance Guidance for orthopedic non-spinal bone screws (i.e. <i>“Orthopedic Non-Spinal Metallic Bone Screws and Washers - Performance Criteria for Safety and Performance Based Pathway (December 2020)”</i>) demonstrated the subject screws met the performance criteria for driving torque. <p>Additionally, Magnetic Resonance compatibility testing has been performed to establish MR Conditional parameters for the subject VOLT™ Mini Fragment and VOLT™ Small Fragment Plating Systems.</p>
Clinical Performance Data	Clinical testing was not necessary for the determination of substantial equivalence.
Substantial Equivalence	<p>The subject devices have the same intended use and similar indications as their predicate devices. Additionally, the subject devices are an iteration on prior DePuy Synthes plating systems and are therefore similar in terms of design, material, and fundamental technology.</p> <p>The non-clinical performance data and analytic evaluations included in this premarket notification demonstrate that any differences in technological characteristics of the subject device compared to the predicate device do not raise any new questions of safety and effectiveness. The proposed devices are at least as safe and effective as the predicate devices.</p>
Conclusion	It is concluded that the information provided demonstrate the substantial equivalence of the subject devices to their predicate devices.